## K道克电子制造(南通)有限公司 LITTLE DOCTOR ELECTRONIC (NANTONG) CO., LTD. No. 8 Tongxing Road Economic and Technical Development Area Nantong Jiangsu P.R. of China

| Declaration of Conformity  |   |
|--|---|
| Manufacturer:  | Little Doctor Electronic (Nantong) Co., Ltd.<br>No. 8, Tongxing Road, Economic and Technical Development<br>Area, 226010 Nantong, Jiangsu, People's Republic of China |
| Facility:  | Little Doctor Electronic (Nantong) Co., Ltd.<br>No. 8, Tongxing Road, Economic and Technical Development<br>Area, 226010 Nantong, Jiangsu, People's Republic of China |
| EU Representative:   | Little Doctor Europe Sp. z o. o.<br>Zawila Str. 57G, 30-390, Krakow, Poland   |
| Product name:  | Aneroid Sphygmomanometers   |
| Model number:  | LD-100  |
| UMDNS Code   | 16156   |
| Classification   | Class I with measuring function (MDR, Annex VIII Rule 1)  |
| UDI-DI   | 697368616001X-697368616999X   |
| Conforming Assessment Route  | Annex IX  |
| We herewith declare that the above mentioned products meet the provisions of the following EC Council Regulation and Standards. All supporting documentations are retained under the premises of the manufacturer. We are exclusively responsible for the Declaration of Conformity. |   |
| General applicable regulation:   | Medical Device regulation: COUNCIL REGULATION (EU)2017/745  |
| Notified Body:   | TÜV SÜD Product Service GmbH,<br>Certification Body, Ridlerstraße 65, 80339 Munich, Germany   |
| Identification №:  | 0123  |
| EC Certificate:  | G2M 071993 0016 Rev.00  |
| Expiration date of the Certificate   | 2024-05-26  |
| Data CE mark was affixed:  | 2006-07   |
| Place:   | People's Republic e China   |
| Date:  | 2021-06-15  |
| Signature:   | 」た 空間 👗 👗 🦉   |
| Name:  | GAO JIAWEN  |
| Position:  | Director  |